

NOTICE

Sealed quotations are invited by the Principal, J. K. College, Purulia from bonafied vendors for **Supply of Digital pH Meter for Chemistry Laboratory** with the following specifications **within 16.09.2025:**

Specifications:

1. 1 High quality precision manual instrument
2. Modes: pH, mV, Temperature
3. pH range: 0.00 - 14.00
4. pH Resolution: 0.01 pH
5. pH Accuracy: ± 0.01 pH
6. pH Buffer option: USA (pH 4.01, 7.00, 10.01)
7. Automatic Buffer recognition
8. mV Range: 0 to ± 1999 mV
9. mV Resolution: 0.01 mV
10. mV Accuracy: ± 0.01 mV
11. Temperature Range: 0°C to 100°C
12. USB Interface
13. LCD Display
14. Memory: up to 50 readings
15. With Reset function.

Please visit <https://www.jkcprl.ac.in/quotation-invitation.php>

For details please see our website **www.jkcprl.ac.in.**

The authority reserves the right to reject or accept any or all quotations without assigning any reasons what so ever.

Principal
J. K. College, Purulia
P. O. & Dist. – Purulia – 723101

Memo No. JKC / Q-9/2025-26 **Dated: 04. 09.2025**

Copy forwarded for information and wide circulation to:-

1. DM, Purulia
2. SDO - (Sadar/East)
3. Zillaparishad, Purulia
4. Head Post Office, Purulia
5. College website & College Notice Board.

Principal
J. K. College, Purulia
P. O. & Dist. – Purulia – 723101